

<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date issued: 21-Jul-2016

FR01S11668175-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process
in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA

Le Coporl

9, avenue ALbert II

MONACO

98000

MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	225/PLANER - 15/07/2016
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	194773
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	E160604

Irradiation Data

Date and Time of Irradiation:	21-JUL-2016 08:44
Calculated Minimum Dose kGy:	27.2
Calculated Maximum Dose kGy:	38.0

Irradiation Release Authorised By Synergy Health plc

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR69 343 092 540