


<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Nov-2017

FR01S12017571-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Coporl
9, avenue ALBERT II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023707
Customer Reference Number:	267/PLANER - 20/11/2017
Product Description:	ESSAI - PROMEPLA
Quantity Received:	7
Customer Unit Lot/Batch Number:	G000575
Other Process Details:	Customer required doses : 25.0 - 45.0 kGy

Irradiation Data

Date and Time of Irradiation:	27-NOV-2017 02:58
Calculated Minimum Dose kGy:	28.6
Calculated Maximum Dose kGy:	39.9

GR062471 / 1
Qty : 3360
RoHS: ?
2019-06-05
LOT : GR062471/001

PN038870/1
CN101568-1
CN101568-1
QS / BOTTLE

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1