



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 18-Apr-2017

FR01S11853854-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process
in accordance with:

EN ISO 11137-1:2015 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA

Le Copori

9, avenue ALbert II

MONACO

98000

MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	244/PLANER - 11/04/2017
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	194773
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	E170306

Irradiation Data

Date and Time of Irradiation:	17-APR-2017 02:54
Calculated Minimum Dose kGy:	28.2
Calculated Maximum Dose kGy:	37.4

Irradiation Release Authorised By Synergy Health plc

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1