



<http://www.synergyhealthplc.com>

Certificate of Irradiation

Date Issued: 12-May-2014

FR01S11157978-1-1

This is to certify that Synergy Health Marseille has where appropriate delivered an irradiation process in accordance with:

EN ISO 11137-1:2006 Sterilisation of Health Care Products

EN ISO 9001:2008 Quality Management System

EN ISO 13485:2012 Quality System - Medical Devices

PROMEPLA

Le Copori

9, avenue ALbert II

MONACO

98000

MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023710
Customer Reference Number:	128/PLANER - 05/05/2014
Product Description:	496604P - DEVICE OOG00199SR148
Validation Reference:	169936
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	42.0
Customer Unit Lot/Batch Number:	E140344

Irradiation Data

Date and Time of Irradiation:	12-MAY-2014 09:23
Calculated Minimum Dose kGy:	28.1
Calculated Maximum Dose kGy:	40.3

Irradiation Release Authorised By Synergy Health plc

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1