

<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Jun-2019

FR01S12290898-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Coporl
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	336/PLANER
Product Description:	496618P - DEVICE 00G00199SR148
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G003077

Irradiation Data

Date and Time of Irradiation:	25-JUN-2019 04:31
Calculated Minimum Dose kGy:	27.5
Calculated Maximum Dose kGy:	41.4

GR067682 / 1

Qty : 3336

Rohs : ?



2019-08-01

LOT : GR067682/001

PN042510/1

CN101568-1

CN101568-1

QS / BOTTLE

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1

Détail de la commande: 336

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: <i>PLANER PLC</i>					
OOG00199SR148		480	10	G003077	R190043
OOG00199SR148		480	10	G003077	R190044
OOG00199SR148		456	10	G003077	R190045
OOG00199SR148		480	10	G003077	R190046
OOG00199SR148		480	10	G003077	R190047
OOG00199SR148		480	10	G003077	R190048
OOG00199SR148		480	10	G003077	R190049
		3336	70		