



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 05-Nov-2019

FR01S12355061-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Coporl
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	351/PLANER - 31/10/2019
Product Description:	496618P - PRODUITS CULTURE CELLULAIRE
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G003811

Irradiation Data

Date and Time of Irradiation:	05-NOV-2019 10:39
Calculated Minimum Dose kGy:	26.4
Calculated Maximum Dose kGy:	37.7

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1

Détail de la commande: 351

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: PLANER PLC					
OOG00199SR148		480	10	G003811	R190090
OOG00199SR148		480	10	G003811	R190091
OOG00199SR148		480	10	G003811	R190092
OOG00199SR148		480	10	G003811	R190093
OOG00199SR148		480	10	G003811	R190094
OOG00199SR148		480	10	G003811	R190095
OOG00199SR148		480	10	G003811	R190096
		3360	70		