



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 05-Mar-2019

FR01S12236144-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	320/PLANER PLC - 11/02/2019
Product Description:	496618P - DEVICE OOG00199SR148
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G002404

Irradiation Data

Date and Time of Irradiation:	05-MAR-2019 12:16
Calculated Minimum Dose kGy:	26.5
Calculated Maximum Dose kGy:	38.8

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Page 1 of 1

Détail de la commande: 320

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: PLANER PLC					
OOG00199SR148		480	10	G002404	R190007
OOG00199SR148		480	10	G002404	R190008
OOG00199SR148		480	10	G002404	R190009
OOG00199SR148		480	10	G002404	R190010
OOG00199SR148		480	10	G002404	R190011
OOG00199SR148		480	10	G002404	R190012
OOG00199SR148		480	10	G002404	R190013
		3360	70		