



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 19-Dec-2018

FR01S12211972-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
EN ISO 13485 Quality System - Medical Devices

PROMEPLA  
Le Coporl  
9 avenue ALbert II  
MONACO  
98000  
MONACO

### Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	313/PLANER - 11/12/2018
Product Description:	496618P - DEVICE OOG00199SR148
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G002256

### Irradiation Data

Date and Time of Irradiation:	17-DEC-2018 11:25
Calculated Minimum Dose kGy:	26.5
Calculated Maximum Dose kGy:	37.8

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE  
N° TVA: FR59 343 092 540