

## CERTIFICATE OF TREATMENT BY GAMMA RADIATION

WE UNDERSIGNED :

**Synergy Health S.A.S.**

MIN 712 - Les ARNAVAUX

13323 MARSEILLE CEDEX 14 - FRANCE



**CERTIFY THAT WE TREATED BY GAMMA RADIATION ACCORDING TO THE SPECIFIC CUSTOMER'S REQUIREMENTS AND TO :**

- the specification of treatment # 496604P
- the specifications of ISO 13485 and ISO 11137-1 standards
- the results of the dose mapping # 169936 of 07/09/2011

**THE FOLLOWING PRODUCTS : (according to the customer's indication)**

**Customer** : PROMEPLA

**Product** : DEVICE OOG00199SR148

**Customer's reference** : ORDER # 98/PLANER  
OF 20/03/2013

**Quantity** : 2 PALLETS

**Irradiation date** : 2013.03.24

**Irradiation dose** : 28.5 kGy to 39.7 kGy

**Irradiation batch number** : 18981601

The control of the applied radiation dose is done by Synergy Health using Red Perspex dosimeters calibrated by the English National Physical Laboratory.

**Synergy Health Marseille,**

**A. NOEL**  
Process Control Officer

A handwritten signature in blue ink, appearing to read 'A. Noel', written over a faint grid.

Certificate # 18981601 / 1

**S. LE GONIDEC**  
Quality Assistant

A handwritten signature in blue ink, appearing to read 'S. Le Gonidec', written over a faint grid.