



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 06-Aug-2020

FR01S12466553-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	373/PLANER - 30/07/2020
Product Description:	496618P - PRODUITS CULTURE CELLULAIRE
Validation Reference:	S12029202
Quantity Received:	7
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G004517

Irradiation Data

Date and Time of Irradiation:	05-AUG-2020 10:20
Calculated Minimum Dose kGy:	27.6
Calculated Maximum Dose kGy:	41.4

GR070996 / 1

Qty : 3360

Rohs: ?



2020-09-29

LOT : GR070996/001

PN044558/1

CN101568-1

CN101568-1

QS / BOTTLE

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Détail de la commande: 373

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: PLANER PLC					
OOG00199SR148		480	10	G004517	R200044
OOG00199SR148		480	10	G004517	R200045
OOG00199SR148		480	10	G004517	R200046
OOG00199SR148		480	10	G004517	R200047
OOG00199SR148		480	10	G004517	R200048
OOG00199SR148		480	10	G004517	R200049
OOG00199SR148		480	10	G004517	R200050
		3360	70		