



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 13-Jul-2020

FR01S12457829-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
EN ISO 13485 Quality System - Medical Devices

PROMEPLA  
Le Copori  
9 avenue ALbert II  
MONACO  
98000  
MONACO

---

### Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1127465
Customer Reference Number:	370/PLANER PLC - 07/07/2020
Product Description:	496618P - PRODUITS CULTURE CELLULAIRE
Validation Reference:	S12029202
Quantity Received:	3
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G004638-G004639-G004640

---

### Irradiation Data

Date and Time of Irradiation:	13-JUL-2020 07:39
Calculated Minimum Dose kGy:	28.4
Calculated Maximum Dose kGy:	39.6

---

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

---

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE  
N° TVA: FR59 343 092 540

Détail de la commande: 370

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
<b>CLIENT: PLANER PLC</b>					
OOG00799SR148		460	10	G004639	R200039
OOG00799SR148		460	10	G004638	R200040
OOG00799SR148		460	10	G004640	R200041
		<b>1380</b>	<b>30</b>		