



<http://www.steris-ast.com>

## Certificate of Irradiation

Date Issued: 22-Apr-2021

FR01S12552872-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products  
EN ISO 13485 Quality System - Medical Devices

PROMEPLA  
Le Copori  
9 avenue ALbert II  
MONACO  
98000  
MONACO

---

### Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1134451
Customer Reference Number:	409/PLANER PLC - 25/03/2021
Product Description:	496633P - OOG00799SR148
Validation Reference:	S12538885
Quantity Received:	6
Customer Minimum Specification kGy:	25.0
Customer Maximum Specification kGy:	45.0
Customer Unit Lot/Batch Number:	G005194+G005721

---

### Irradiation Data

Date and Time of Irradiation:	22-APR-2021 06:13
Calculated Minimum Dose kGy:	26.1
Calculated Maximum Dose kGy:	39.8

---

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

---

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE

N° TVA: FR59 343 092 540

Détail de la commande: 409

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
<b>CLIENT: PLANER PLC</b>					
OOG00799SR148		480	10	G005194	R210030
OOG00799SR148		480	10	G005721	R210031
OOG00799SR148		480	10	G005721	R210032
OOG00799SR148		480	10	G005721	R210033
OOG00799SR148		480	10	G005721	R210034
OOG00799SR148		480	10	G005721	R210035
		2880	60		