

Détail de la commande: 293

Référence	Ref client	Quantité	Nb carton	Num lot	N° palette
CLIENT: PLANER PLC					
OOG00499SR148		96	1	P180302	R180036
OOG00499SR148		1824	19	N001670	R180036
		1920	20		

R180036 → samples
 for
 dose
 mapping.
 not for human use.
 A. Bagella



<http://www.steris-ast.com>

Certificate of Irradiation

Date Issued: 28-Jun-2018

FRO1S12124416-1-1

This is to certify that Synergy Health Marseille, a STERIS Company has where appropriate delivered an irradiation process in accordance with the current certified standards:

EN ISO 11137-1 Sterilisation of Health Care Products
EN ISO 13485 Quality System - Medical Devices

PROMEPLA
Le Copori
9 avenue ALbert II
MONACO
98000
MONACO

Order Information

Account Number:	104501
Synergy Health Sales Part Reference:	1023707
Customer Reference Number:	165A-18AB - 14/06/2018
Product Description:	ESSAI - PROMEPLA
Quantity Received:	1
Customer Unit Lot/Batch Number:	N001670
Other Process Details:	Customer required doses : 20.0 - 45.0 kGy This certificate cancels and replaces the last version.

Irradiation Data

Date and Time of Irradiation:	24-JUN-2018 23:13
Calculated Minimum Dose kGy:	25.5
Calculated Maximum Dose kGy:	33.5

Irradiation Release Authorised By Synergy Health Marseille SAS, a STERIS Company

Processing Site: M.I.N. 712 - Arnavaux, , Marseille Cedex 14, 13323 Phone No: +33 (0) 4 91 214 214

Registered Office: M.I.N. 712 - Arnavaux, 13323 Marseille Cedex 14, FRANCE
N° TVA: FR59 343 092 540